

Designation :	Id-Fungi Plates	BATCH N° :
Presentation :	Petri dishes 90 mm	
Production date :	02/07/2024	
Expiration date :	15/01/2025	38027
Control date :	03/07/2024	

PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance	Translucent Beige	CONFORM
pH	4,5 to 6,0 (indicative value)	pH=5,4
WIDTH (mm)	>3mm	CONFORM
MICROBIOLOGICAL STATE		
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM
MICROBIOLOGICAL ACTIVITY		
Inoculum < 100 cfu Candida albicans ATCC 10231 / WDCM 00054	After 48h-72h at 20-25°C Growth	CONFORM
Inoculum > 100 cfu Escherichia coli ATCC 8739 / WDCM 00012	After 48h at 20-25°C Inhibition	CONFORM
Inoculum : NA Aspergillus brasiliensis ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C Growth	CONFORM
Trichophyton rubrum IP 1464.83	5 to 10 days at 20-25°C Growth	CONFORM

AQ CONCLUSION	CONFORM
	BATCH ACCEPTED

Date :

23 / 07 / 24

Responsable Qualité:



CONIDIA CONIPHY
2 rue en Chuel
PA en Chuel
69650 Quincieux – France