

Product name : Id-Fungi Plates Plus BATCH N° :
Presentation : Petri dishes 90 mm
Production date : 12/02/2025
Expiration date : 28/08/2025 **39617**
Control date : 13/02/2025

PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance pH	Beige indicative value	CONFORM pH=5.6

MICROBIOLOGICAL STATE	SPECIFICATIONS	RESULTS
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM

MICROBIOLOGICAL ACTIVITY	SPECIFICATIONS	RESULTS
Inoculum > 100 CFU <i>Escherichia coli</i> ATCC 8739 / WDCM 00012	After 48H-72H at 20-25°C INHIBITION	CONFORM
Inoculum : N/A <i>Aspergillus brasiliensis</i> ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C INHIBITION	CONFORM
Inoculum : N/A <i>Trichophyton rubrum</i> IP 1464.83	5 to 10 days at 20-25°C GROWTH	CONFORM

CONCLUSION

CONFORM

NOT CONFORM

☒ BATCH ACCEPTED

☐ BATCH NOT ACCEPTED

Date :

03/03/2025

AQ:



CONIDIA CONIPHY
2 rue en Chuel
PA en Chuel
69650 Quincieux – France



[Signature]
NB