

|                   |                    |              |
|-------------------|--------------------|--------------|
| NOM Produit :     | Id-Fungi Plates    | BATCH N° :   |
| Presentation :    | Petri dishes 90 mm |              |
| Production date : | 24/06/2025         |              |
| Expiration date : | 07/01/2026         | <b>40600</b> |
| Control date :    | 25/06/2025         |              |

| PHYSICAL-CHEMICAL TESTS | SPECIFICATIONS            | RESULTS        |
|-------------------------|---------------------------|----------------|
| Appearance<br>pH        | Beige<br>indicative value | CONFORM<br>5.6 |

| MICROBIOLOGICAL STATE                                     | SPECIFICATIONS                                 | RESULTS |
|-----------------------------------------------------------|------------------------------------------------|---------|
| Direct incubation<br>before and after the membrane set up | Sterile after 7 days<br>at 20-25°C and 30-35°C | CONFORM |

| MICROBIOLOGICAL ACTIVITY                                                                | SPECIFICATIONS                     | RESULTS |
|-----------------------------------------------------------------------------------------|------------------------------------|---------|
| <b>Inoculum &lt; 100 CFU</b><br><br><i>Candida albicans</i><br>ATCC 10231 / WDCM 00054  | After 48H-72H at 20-25°C<br>GROWTH | CONFORM |
| <b>Inoculum &gt; 100 CFU</b><br><br><i>Escherichia coli</i><br>ATCC 8739 / WDCM 00012   | After 48H at 20-25°C<br>INHIBITION | CONFORM |
| <b>Inoculum : N/A</b><br><br><i>Aspergillus brasiliensis</i><br>ATCC 16404 / WDCM 00053 | Until 5 days at 20-25°C<br>GROWTH  | CONFORM |
| <i>Trichophyton rubrum</i><br>IP 1464.83                                                | 5 to 10 days at 20-25°C<br>GROWTH  | CONFORM |

|                                                                                                   |                |                    |
|---------------------------------------------------------------------------------------------------|----------------|--------------------|
| <b>CONCLUSION</b>                                                                                 | <b>CONFORM</b> | <b>NOT CONFORM</b> |
| <input checked="" type="checkbox"/> BATCH ACCEPTED<br><input type="checkbox"/> BATCH NOT ACCEPTED |                |                    |

Date : 09/07/2025

AQ:



CONIDIA CONIPHY  
2 rue en Chuel  
PA en Chuel  
69650 Quincieux – France

