

| Product name :  | Id-Fungi Plates Plus                           | BATCH N° :         |
|---|--|--------------------|
| Presentation :  | Petri dishes 90 mm                             |                    |
| Production date :   | 10/02/2026                                     |                    |
| Expiration date :   | 26/08/2026                                     | <b>41940</b>       |
| Control date :  | 12/02/2026                                     |                    |
| PHYSICAL-CHEMICAL TESTS   | SPECIFICATIONS                                 | RESULTS            |
| Appearance<br>pH  | Beige<br>indicative value                      | CONFORM<br>pH=5.2  |
| MICROBIOLOGICAL STATE   | SPECIFICATIONS                                 | RESULTS            |
| Direct incubation<br>before and after the membrane set up   | Sterile after 7 days<br>at 20-25°C and 30-35°C | CONFORM            |
| MICROBIOLOGICAL ACTIVITY  | SPECIFICATIONS                                 | RESULTS            |
| <b>Inoculum &gt; 100 CFU</b><br><br><i>Escherichia coli</i><br>ATCC 8739 / WDCM 00012             | After 48H-72H at 20-25°C<br>INHIBITION         | CONFORM            |
| <b>Inoculum : N/A</b><br><br><i>Aspergillus brasiliensis</i><br>ATCC 16404 / WDCM 00053           | Until 5 days at 20-25°C<br>INHIBITION          | CONFORM            |
| <b>Inoculum : N/A</b><br><br><i>Trichophyton rubrum</i><br>IP 1464.83                             | 5 to 10 days at 20-25°C<br>GROWTH              | CONFORM            |
| <b>CONCLUSION</b>   |  | <b>CONFORM</b>     |
|   |  | <b>NOT CONFORM</b> |
| <input checked="" type="checkbox"/> BATCH ACCEPTED<br><input type="checkbox"/> BATCH NOT ACCEPTED |  |                    |

Date :

25/02/2026

AQ:



CONIDIA CONIPHY  
2 rue en Chuel  
PA en Chuel  
69650 Quincieux – France


