

	FORMULAIRE:		Diffusion : Réaliser Produits
	CERTIFICATE OF ANALYSES / CERTIFICAT D'ANALYSES		
	Approuvé par le pilote : CD	Vérifié par Responsable Qualité : NB	MAJ : 03/03/2025
			Page : 1/1

Product name :	Id-Fungi Plates Plus	BATCH N° :	
Presentation :	Petri dishes 90 mm		
Production date :	08/07/2026		
Expiration date :	21/01/2027		43011
Control date :	09/07/2026		

PHYSICAL-CHEMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance pH	Beige indicative value	CONFORM pH=5.5

MICROBIOLOGICAL STATE	SPECIFICATIONS	RESULTS
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM

MICROBIOLOGICAL ACTIVITY	SPECIFICATIONS	RESULTS
Inoculum > 100 CFU <i>Escherichia coli</i> ATCC 8739 / WDCM 00012	After 48H-72H at 20-25°C INHIBITION	CONFORM
Inoculum : N/A <i>Aspergillus brasiliensis</i> ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C INHIBITION	CONFORM
Inoculum : N/A <i>Trichophyton rubrum</i> IP 1464.83	5 to 10 days at 20-25°C GROWTH	CONFORM

CONCLUSION	CONFORM	NOT CONFORM
------------	---------	-------------

☒ BATCH ACCEPTED
☐ BATCH NOT ACCEPTED

Date :

21/07/2026

AQ:




CONIDIA CONIPHY
 2 rue en Chuel
 PA en Chuel
 69650 Quincieux – France

