CONIDIA CONIPHY	FORMULAIRE:	Diffusion : Réaliser Produits
	CERTIFICATE OF ANALYSES / CERTIFICAT D'ANALYSES	
	Approuvé par le pilote : CD / Vérifié par Responsable Qualité : NB	MAJ : 03/03/2025
		Page : 1/1

Product name :	Id-Fungi Plates Plus	BATCH N° :	
Presentation :	Petri dishes 90 mm		
Production date :	13/10/2026		
Expiration date :	28/04/2027		**43747**
Control date :	14/10/2026		

PHYSICAL-CHIMICAL TESTS	SPECIFICATIONS	RESULTS
Appearance	Beige	CONFORM
pH	indicative value	5,3
MICROBIOLOGICAL STATE	**SPECIFICATIONS**	**RESULTS**
Direct incubation before and after the membrane set up	Sterile after 7 days at 20-25°C and 30-35°C	CONFORM
MICROBIOLOGICAL ACTIVITY	**SPECIFICATIONS**	**RESULTS**
Inoculum > 100 CFU *Escherichia coli* ATCC 8739 / WDCM 00012	After 48H-72H at 20-25°C INHIBITION	CONFORM
Inoculum : N/A *Aspergillus brasiliensis* ATCC 16404 / WDCM 00053	Until 5 days at 20-25°C INHIBITION	CONFORM
Inoculum : N/A *Trichophyton rubrum* IP 1464.83	5 to 10 days at 20-25°C GROWTH	CONFORM
CONCLUSION	**CONFORM**	~~**NOT CONFORM**~~

☑ **BATCH ACCEPTED**
☐ **BATCH NOT ACCEPTED**

Date : 26/10/2026

AQ:

CONIDIA CONIPHY
2 rue en Chuel
PA en Chuel
69650 Quincieux – France

CE IVD